#### QUICKTIP

# Troubleshooting Guide

Neuro Processor Technology

#### Patient's Own Voice

#### Voice Sounds

- In a barrel/tunnel
- **Echoes**
- Hollow
- Like they have a cold/ears plugged

#### **Pro Fit Adjustments**

- Decrease gain using the Occlusion Control
- Decrease Low Frequency Gain
- Decrease Moderate Gain at 1000 Hz and/or 1500 Hz

#### **Voice Sounds**

Muffled

#### **Pro Fit Adjustments**

- Increase Loud Gain
- Increase Maximum Output
- Increase High Frequency Gain
- Decrease Low Frequency Gain

#### **Voice Sounds**

- Distorted
- Crackles
- Unnatural/like a megaphone

- Increase Moderate Gain at 1000 Hz and/or 1500 Hz

#### **Pro Fit Adjustments**

- Decrease Moderate Gain at 1000 Hz and/or 1500 Hz
- Decrease Loud Gain
- Decrease Overall Output

#### **Other Considerations**

- Occlusion may be due to the physical presence of the hearing aid and not because of amplification; to test, turn off the hearing aid and have the patient speak
  - 1. Report persists—issue is occlusion; address with acoustic modifications
    - Enlarge vent diameter
    - Shorten and/or taper canal
    - Remake hearing aid or earmold with different canal length
  - 2. Report resolved—issue is amplification; address with response adjustments

#### **Other Considerations**

- Occlusion may be due to the physical presence of the hearing aid and not because of amplification; to test, turn off the hearing aid and have the patient speak
  - 1. Report persists—issue is occlusion; address with acoustic modifications
    - Enlarge vent diameter
    - Shorten and/or taper canal
    - Remake hearing aid or earmold with different canal length
  - 2. Report resolved issue is amplification; address with response adjustments

#### Other Considerations

 If decreasing Overall Output worsens sound quality, consider increasing Overall Output

### Hearing in Noise

#### **Patient has Difficulty**

 Understanding speech in background noise

#### **Pro Fit Adjustments**

- Verify Adaptive Directionality is enabled via Sound Manager screen
- Consider enabling a Fixed Directional microphone response
- Increase Speech in Noise Control via Sound Manager screen
- Increase Overall Gain at 1000 Hz and/or 1500 Hz, then higher frequency gain
- Decrease Soft Low Frequency Gain

#### **Pro Fit Adjustments**

- Increase Overall Gain at 1000 Hz and/or 1500 Hz
- Increase Overall Soft Gain
- Decrease Speech in Noise setting via the Sound Manager screen

#### **Patient Reports**

**Patient Hears** 

Voices at a distance

better than nearby

- Low tolerance for noise
- Background noise too loud

#### **Pro Fit Adjustments**

- Decrease Overall Output
- Verify Adaptive Directionality is enabled via Sound Manager screen
- Increase Speech in Noise setting via the Sound Manager screen
- Increase Transients setting via Sound Manager screen

#### Other Considerations

- If device does not have directional microphones, consider recommending a directional device
- Consider enabling Edge Mode+ via the User Control screen
- Consider a StarLink Remote Control 2.0 and set the Favorite Button to Edge Mode+
- Consider StarLink remote microphone options to improve signal-to-noise ratio
- Consider turning Speech in Noise off for severe-to-profound hearing loss

#### **Other Considerations**

- If device does not have directional microphones, consider recommending a directional device
- Consider enabling Edge Mode+ via the User Control screen
- Consider a StarLink Remote Control 2.0 and set the Favorite Button to Edge Mode+
- Consider StarLink remote microphone options to improve signal-to-noise ratio

#### Other Considerations

- If device does not have directional microphones, consider recommending a directional device
- Consider enabling Edge Mode+ via the User Control screen
- Consider a StarLink Remote Control 2.0 and set the Favorite Button to Edge Mode+
- Consider StarLink remote microphone options to improve signal-to-noise ratio

## Intelligibility

#### Reports

- I hear better without my hearing aids
- Speech is unclear/unnatural
- Speech in quiet is not clear

Muffled even when in guiet

TV/Radio is not clear

#### **Pro Fit Adjustments**

- Increase Overall Gain at 1000 Hz and/or 1500 Hz
- Decrease Speech in Noise setting via the Sound Manager screen
- Decrease Speech in Noise setting via the Sound Manager screen
- Increase Low Frequency Gain for Streamed program

#### Speech Sounds Pro Fit Adjustments

- Decrease Quiet setting via the Sound Manager screen
- Increase Soft and Moderate Gain

#### Other Considerations

- Consider a customized program via their hearing aid mobile app
- Consider adding a StarLink TV Streamer
- Consider enabling Edge Mode+ via the User Control screen
- Consider StarLink Remote Microphone options to improve signal-to-noise ratio
- May need to counsel on fact that poor speech clarity may be due to poor speech discrimination
- Consider a StarLink Remote Control 2.0 and set the Favorite Button to Edge Mode+

#### **Other Considerations**

- Quiet adjusts expansion and low-level noise reduction to ensure the hearing aids are quiet in quiet environments
- Consider enabling Edge Mode+ via the User Control screen

### Streamed Input

#### **Reports**

Streamed input doesn't have enough bass

#### **Reports**

Streamed input is not loud enough

#### **Reports**

External environment is louder than the streamed signal

#### **Pro Fit Adjustments**

- Enable Bass Boost via the Accessories screen
- Increase Gain for Low Frequencies
- Increase Output for Low Frequencies

#### **Pro Fit Adjustments**

- Enable Overall Boost via the Accessories screen
- Enable Treble Boost for phone calls via the Accessories screen
- Enable Bass Boost via the Accessories screen
- Enable a program with increased low frequency gain for the patient to switch to before they begin streaming
- Enable a Music program for the patient to switch to before they begin streaming

#### **Pro Fit Adjustments**

Mute the hearing aid microphones

#### **Other Considerations**

Consider a customized program via the hearing aid mobile app

#### **Other Considerations**

 Consider setting the Hearing Aid Mic While Streaming setting to Off via the Accessories screen or the hearing aid mobile app

#### **Other Considerations**

 Consider setting the Hearing Aid Mic While Streaming setting to Off via the Accessories screen or the hearing aid mobile app.

### Loudness

#### **Overall Too Loud**

- Voices too loud
- All sounds too loud
- Harsh/too loud

#### **Pro Fit Adjustments**

- Change Experience Level to provide less gain (3 to 2 or 2 to 1)
- Decrease Gain using Occlusion Control
- Decrease High Frequency Loud Gain

#### **Loudness Comfort**

- Sounds are painful
- Clattering dishes too loud
- Running water
- Other environmental sounds too loud

#### **Overall Too Soft**

- Voices too soft
- All sounds too soft
- Hearing aids too soft

- Decrease Overall Gain above 1000 Hz

- **Pro Fit Adjustments**
- Increase Transients setting via the Sound Manager screen
- Decrease High Frequency Loud Gain
- Decrease Overall Output
- Decrease Overall Loud Gain
- Increase Machine Noise setting via the Sound Manager screen

#### **Pro Fit Adjustments**

- Increase Overall Gain
- Increase Overall Output
- Increase Overall Soft Gain
- Increase Overall Moderate Gain
- Increase Low Frequency Overall Gain
- Decrease Quiet setting via the Sound Manager screen

#### **Other Considerations**

- May need to start with lower gain settings than the prescriptive target recommends
- Patient may be unaccustomed to amplification or may be accustomed to lower gain devices
- May need to consider a different fitting formula
- Compression ratios are increased as the curves move closer together; decreased as the curves move farther apart.
- Consider enabling Edge Mode+ via the User Control screen.

#### **Other Considerations**

- Ensure Best Fit is using e-STAT 2.0 fitting formula
- Enter pure tone UCLs for at least 500 Hz and 3000 Hz to personalize and help optimize the output settings
- Utilize Speech Mapping to identify frequencies causing discomfort
- Compression Ratios are increased as the curves move closer together; decreased as the curves move farther apart

#### Other Considerations

- Utilize Speech Mapping to verify audibility
- Patient may not perceive the aid as being loud enough depending on previous hearing aid experience
- Compression Ratios are increased as the curves move closer together; decreased as the curves move farther apart
- Quiet adjusts expansion and low-level noise reduction to ensure the hearing aids are quiet in quiet environments

## **Sound Quality**

#### Noisy

- Hearing aids are noisy
- Refrigerator hum too loud
- Hearing aids are noisy in quiet environments

### Pumping

- Hearing aids cut in and out
- Hearing aids cut in and out when patient speaks
- Loud sounds fade in and out

#### **Shutting Down**

- Hearing aids shut down with loud sounds
- Hearing aids cut out when patient speaks
- Loud sounds fade in and out

#### **Transient Sounds are**

Bothersome

#### **Transient Sounds are**

Too Soft/Unnatural

#### Sounds are

- Hollow
- Muffled

#### **Pro Fit Adjustments**

- Increase Quiet setting via Sound Manager screen
- Decrease Soft Gain at 750 Hz and below
- Decrease Overall Soft Gain

#### **Pro Fit Adjustments**

- Increase Overall Loud Gain
- Decrease Compression Ratios
- Decrease Machine Noise setting via Sound Manager screen
- Decrease Consonant Brightness

#### **Pro Fit Adjustments**

- Decrease Transients setting via the Sound Manager screen
- Decrease Compression Ratios
- Increase Overall Output/MPO
- Increase Overall Gain
- Increase Overall Soft Gain
- Increase Overall Loud Gain

### **Pro Fit Adjustments**

Increase Transients setting via Sound Manager screen

#### **Pro Fit Adjustments**

Decrease Transients setting via Sound Manager screen

#### **Pro Fit Adjustments**

- Decrease Loud Gain at 500 Hz and 750 Hz
- Increase Moderate Gain at 1000 Hz and/or 1500 Hz
- Increase Moderate High Frequency Gain

#### Other Considerations

 Quiet adjusts expansion and low-level noise reduction to ensure the hearing aids are quiet in quiet environments

#### Other Considerations

 Compression Ratios are increased as the curves move closer together; decreased as the curves move farther apart

#### **Other Considerations**

 Compression Ratios are increased as the curves move closer together; decreased as the curves move farther apart

#### Other Considerations

Consider turning off for severe-to-profound hearing loss

#### **Other Considerations**

Consider turning off for severe-to-profound hearing loss

#### **Other Considerations**

Increase Vent Size and update Acoustic Options to match hearing aid

## Sound Quality (Continued)

#### Sounds are

- Sharp
- Tinny

#### **Pro Fit Adjustments**

- Increase gain between 2000 Hz-4000 Hz, then increase gain at 750 Hz
- Increase Low Frequency Gain
- Decrease Overall Output above 1000 Hz
- Increase Speech in Noise
- Increase Compression
- Change Experience Level to provide less gain (3 to 2 or 2 to 1)
- Consider enabling the Automatic feature within Experience Manager

#### **Other Considerations**

- Utilize Speech Mapping or Verify Comfort to identify areas of sharpness
- Compression Ratios are increased as the curves move closer together: decreased as the curves move farther apart
- Consider Best Fit using a different fitting formula
- Patient's auditory perception may be distorted due to long-standing high-frequency hearing loss; counseling is key

### Music

#### **Music Sounds**

Too tinny in the Music Program

## **Pro Fit Adjustments**

- Decrease treble via QuickFit screen
- Increase bass via QuickFit screen

#### **Music Sounds**

Too much bass in the Music Program

#### **Pro Fit Adjustments**

- Decrease bass via QuickFit screen
- Increase treble via QuickFit screen

#### Other Considerations

- Consider the Fine-Tuning screen for patients who require very discrete frequency-specific adjustments
- Consider use of StarLink streaming accessory
- Consider enabling Edge Mode+ via the User Control screen

#### Other Considerations

- Consider the Fine-Tuning screen for patients who require very discrete frequency-specific adjustments
- Consider use of a StarLink streaming accessory
- Consider enabling Edge Mode+ via the User Control screen

### Feedback

#### **Hearing Aids**

**Hearing Aids** 

Sound warbly with own

voice or other inputs

- Whistle
- Chirp

### **Pro Fit Adjustments**

- Reduce Adaptive Feedback Cancellation Sensitivity (Strong to Subtle, Subtle to Off) via the Feedback Canceller screen
- Reduce Overall Gain

- Initialize feedback cancellation with hearing aid in the ear

#### **Pro Fit Adjustments**

- Initialize feedback cancellation with hearing aid in the ear
- Reduce Adaptive Feedback Cancellation Sensitivity (Strong to Subtle, Subtle to Off) via the Feedback Canceller screen
- Decrease Consonant Brightness

#### Other Considerations

- Manage acoustic options for better fit and positioning of the hearing aid in the ear
- Utilize Speech Mapping to identify feedback peak and decrease gain at peak
- Feedback cancellation needs to be re-initialized any time the acoustic characteristics of the hearing aid are changed (e.g. shell modification, new earmold)

#### **Other Considerations**

- Manage acoustic options for better fit and positioning of the hearing aid in the ear
- Utilize Speech Mapping to identify feedback peak and decrease gain at peak
- Feedback cancellation needs to be re-initialized any time the acoustic characteristics of the hearing aid are changed (e.g. shell modification, new earmold)



### The Starkey logo, Pro Fit, Neuro Processor, e-STAT, StarLink and Edge Mode are trademarks of Starkey Laboratories, Inc. ©2024 Starkey Laboratories Inc. All Rights Reserved. 10/24 FLYR4093-02-EN-SC